





Certificate of Analysis

SALSA® MLPA® Reagent Kit

| | | |
|---|---|-------------|
| Catalogue # | EK1-FAM | |
| Product name | SALSA MLPA Reagent Kit | |
|  | 26.0010 | |
|  | 100 reactions. | |
|  | Expiration date: July 2028. | |
| Quality control specifications | | |
| Reagent kit components | <ul style="list-style-type: none"> - Test with selected quality control probemixes to check for hybridization speed, ligase-65 activity and polymerase activity. Variation of primer peaks between new and old lots is <20%. Variation within the reference sample population and within the sample population is <10% (maximum stdv 0.10). Residual primer% is below 40%. Probe ratio of special testing probes is between 0.8-1.2. No DNA reaction only shows the 64, 70, 76 and 82 nt control fragments. - Primer-dimer-like peak formation is tested on the PCR Primer mix by using a mix of LPOs that have extensive homology with the PCR Primers. The primer-dimer-like peak is lower than threefold the average height of the probe peaks. - Ligase Buffer B conductivity variation between lots is <10%. - Tests performed (not on Ligase Buffer A and Ligase-65) on 5 ng DNA to determine variation of the primer peaks. This variation should be ≤20%. - Stability test (incubation for 1 week at 37°C). Variation of primer peaks between new and old lots is <20%. Variation within the reference sample population and within the sample population is <10% (maximum stdv 0.10). Residual primer % is below 40%. Probe ratio of special testing probes is between 0.8-1.2. No DNA reaction only shows the 64, 70, 76 and 82 nt control fragments. Sloping is <60%. | Test result |
| | PASS | |
| Assembled reagent kit | <ul style="list-style-type: none"> - Test with selected quality control probemix on male and female samples to check for correct fragment profile under normal testing conditions. | PASS |

This lot was certified by MRC Holland on 27 March 2026.

This certificate is a declaration of analysis at the time of the manufacturing process. All assays were run in compliance with manufacturer's instructions for use.

| | |
|---|---|
| More information: www.mrcholland.com | |
|  | MRC Holland BV; Willem Schoutenstraat 1 1057 DL, Amsterdam, The Netherlands |
| E-mail | info@mrcholland.com (information & technical questions) order@mrcholland.com (orders) |
| Phone | +31 888 657 200 |

Implemented changes in the COA

Version 01 – 27 March 2026 (01)
- Not applicable, new document.