

Certificate of Analysis SALSA® Binning DNA SD031

Catalogue #	SD031	
Product name	SALSA Binning DNA SD031	
LOT	S01-0921	
Σ	6 reactions.	
Shipping conditions	Dry ice or cooling elements.	
1	Store upon arrival between -25 °C and -15 °C.	
\subseteq	Expiry date: September 2026, when stored at recommended conditions.	
Purpose	To be used with SALSA MLPA Probemix P140-C1 HBA, SALSA MLPA reagent kits and Coffalyser.Net™ analysis software as described in the MLPA General Protocol and the corresponding probemix product description.	
Quality control specifications	- The mutation-specific probes, as described in the product description, produce a signal at the designed length.	Test result
	- The signal of the mutation-specific probes when tested with wild-type genomic DNA is <5% of the peak height generated when tested with SALSA Binning DNA SD031.	
	 The signal of the mutation-specific probes when tested with SALSA Binning DNA SD031 is at least 15% in peak height of the average probe signal of the 10 neighbouring probes and is not >25% higher than any of the 10 neighbouring probes. 	PASS
	- The 105 nt chromosome Y specific control fragment generates a signal on SALSA Binning DNA SD031 which is similar (±20%) to the signal obtained on wild-type male genomic DNA.	
	 All SALSA MLPA Probemix P140-C1 HBA probes, other than the mutation-specific probes, show normal signals, similar (±20%) to the peak pattern obtained on wild-type genomic DNA. 	

More information: www.mrcholland.com; www.mrcholland.eu		
	MRC Holland bv; Willem Schoutenstraat 1	
	1057 DL, Amsterdam, The Netherlands	
E-mail	info@mrcholland.com (information & technical questions)	
	order@mrcholland.com (orders)	
Phone	+31 888 657 200	



Certificate of Analysis SALSA® Binning DNA SD031 sample pictures




Figure 1. Capillary electrophoresis pattern from SALSA Binning DNA SD031-S01-0921 (approximately 50 ng) analysed with SALSA MLPA Probemix P140 HBA (C1-0919). The location of the Hb Constant Spring mutation-specific probe at 136 nt is indicated. Probe peak heights may vary between different lots of the P140-C1 probemix.








Figure 2. Capillary electrophoresis pattern from a sample of approximately 50 ng human male control DNA analysed with SALSA MLPA Probemix P140 HBA (C1-0919). Probe peak heights may vary between different lots of the P140-C1 probemix.

This lot was certified by MRC Holland on 05 October 2021.

This certificate is a declaration of analysis at the time of the manufacturing process. All assays were run in compliance with manufacturer's instructions for use.

Implemented changes in the COA

Version 01 - 12 October 2021(05)

- Not applicable, new document.