




Certificate of Analysis

SALSA® MLPA® Probemix P496 KMT2A

Catalogue #	P496-025R, P496-050R, P496-100R	
Product name	Probemix P496 KMT2A	
LOT	A1-1023	
	25, 50, or 100 reactions.	
Shipping conditions	Dry ice or cooling elements.	
	Store upon arrival between -25°C and -15°C.	
	Expiration date: October 2028, when stored at recommended conditions. This product should not be frozen/thawed more than 25 times.	
Purpose	The SALSA MLPA Probemix P496 KMT2A has been developed to determine the DNA copy number of the <i>KMT2A</i> gene, and to detect the presence of <i>ASXL1</i> c.1934dupG mutation. This probemix is designed for use only in combination with SALSA MLPA reagent kits and, SD096 and Coffalyser.Net analysis software as described in the MLPA General Protocol.	
Quality control specifications	<ul style="list-style-type: none"> - Sufficient distance between peaks, absence of extra or shoulder peaks, and completeness of hybridisation of each individual probe, as tested on Applied Biosystems and Beckman/SCIEX GeXP sequencers. - Standard deviation of each individual probe ≤ 0.10, when tested on 23 different DNA samples of healthy individuals, extracted by various methods. - Each individual probe meets reaction-specific criteria when tested on a single DNA sample under various experimental conditions. - No-DNA controls result in only five major peaks shorter than 121 nucleotides (nt): four Q-fragments at 64, 70, 76 and 82 nt, and one peak in the range of 0-40 nt corresponding to the unused portion of the fluorescent PCR primer. Non-specific peaks longer than 121 nt AND with a height <25% of the median of the four Q-fragments are not expected to affect MLPA reactions when sufficient (50-250 ng) sample DNA is used. 	Test result
		PASS

None of the ingredients are derived from humans, animals, or pathogenic bacteria. Based on the concentrations present, none of the ingredients are hazardous as defined by the Hazard Communication Standard. **A Safety Data Sheet (SDS) is not required for these products:** none of the preparations contain dangerous substances (as per Regulation (EC) No 1272/2008 [EU-GHS/CLP] and amendments) at concentrations requiring distribution of an SDS (as per Regulation (EC) No 1272/2008 [EU-GHS/CLP] and 1907/2006 [REACH] and amendments). If spills occur, clean with water and follow appropriate site procedures.

More information: www.mrcholland.com ; www.mrcholland.eu	
	MRC Holland bv; Willem Schoutenstraat 1 1057 DL, Amsterdam, The Netherlands
E-mail	info@mrcholland.com (information & technical questions) order@mrcholland.com (orders)
Phone	+31 888 657 200

Certificate of Analysis

SALSA MLPA Probemix P496-A1 KMT2A sample picture




Figure 1. Capillary electrophoresis pattern from a sample of approximately 50 ng human male control DNA analysed with SALSA MLPA Probemix P496 KMT2A (A1-1023).




Figure 2 Capillary electrophoresis pattern from SALSA Binning DNA SD096-S01 (approximately 50 ng) analysed with SALSA MLPA Probemix P496 KMT2A (A1-1023). The location of the ASXL1 c.1934dupG mutation specific probe at 160 nt is indicated.

This lot was certified by MRC Holland on 13 November 2023.

This certificate is a declaration of analysis at the time of the manufacturing process. All assays were run in compliance with manufacturer's instructions for use.

Implemented changes in the COA
Version 01 – 13 November 2023 (6) - Not applicable, new document.