

Certificate of Analysis SALSA® MLPA® Probemix P383 T-ALL

Catalogue #	P383-025R, P383-050R, P383-100R	
Product name	Probemix P383 T-ALL	
LOT	A2-0923	
Σ	25, 50, or 100 reactions.	
Shipping conditions	Dry ice or cooling elements.	
X	Store upon arrival between -25°C and -15°C.	
	Expiration date: September 2028, when stored at recommended conditions. This product should not be frozen/thawed more than 25 times.	
Purpose	This product has been developed to detect deletions or gains in 13 different chromosomal regions, as described in Tables 1 and 2 of the P383-A2 product description. This probemix is designed for use only in combination with SALSA MLPA reagent kits and Coffalyser.Net analysis software as described in the MLPA General Protocol.	
Quality control specifications	 Sufficient distance between peaks, absence of extra or shoulder peaks, and completeness of hybridisation of each individual probe, as tested on Applied Biosystems and Beckman/SCIEX GeXP sequencers. Standard deviation of each individual probe ≤0.10, when tested on 23 different DNA samples of healthy individuals, extracted by various methods. Each individual probe meets reaction-specific criteria when tested on a single DNA sample under various experimental conditions. No-DNA controls result in only five major peaks shorter than 121 nucleotides (nt): four Q-fragments at 64, 70, 76 and 82 nt, and one peak in the range of 0-40 nt corresponding to the unused portion of the fluorescent PCR primer. Non-specific peaks longer than 121 nt AND with a height <25% of the median of the four Q-fragments are not expected to affect MLPA reactions when sufficient (50-250 ng) sample DNA is used. 	Test result PASS

None of the ingredients are derived from humans, animals, or pathogenic bacteria. Based on the concentrations present, none of the ingredients are hazardous as defined by the Hazard Communication Standard. A Safety Data Sheet (SDS) is not required for these products: none of the preparations contain dangerous substances (as per Regulation (EC) No 1272/2008 [EU-GHS/CLP] and amendments) at concentrations requiring distribution of an SDS (as per Regulation (EC) No 1272/2008 [EU-GHS/CLP] and 1907/2006 [REACH] and amendments). If spills occur, clean with water and follow appropriate site procedures.

More information: www.mrcholland.com; www.mrcholland.eu		
•••	MRC Holland bv; Willem Schoutenstraat 1	
	1057 DL, Amsterdam, The Netherlands	
E-mail	info@mrcholland.com (information & technical questions)	
	order@mrcholland.com (orders)	
Phone	+31 888 657 200	



Certificate of Analysis SALSA MLPA Probemix P383-A2 T-ALL sample picture




Figure 1. Capillary electrophoresis pattern from a sample of approximately 50 ng human male control DNA analysed with SALSA MLPA Probemix P383 T-ALL (A2-0923).




Figure 2 Capillary electrophoresis pattern from a sample of approximately 50 ng human female control DNA analysed with SALSA MLPA Probemix P383 T-ALL (A2-0923).

This lot was certified by MRC Holland on 07 February 2024.

This certificate is a declaration of analysis at the time of the manufacturing process. All assays were run in compliance with manufacturer's instructions for use.

Implemented changes in the COA

Version 01 - 07 February 2024(6)

- Not applicable, new document.