

Certificate of Analysis

SALSA® Reference Selection DNA SD082

Catalogue #	SD082	
Product name	SALSA® Reference Selection DNA SD082 (hereafter: SD082)	
LOT	S01-0126	
Shipping conditions	Dry ice or cooling elements.	
	Expiration date: January 2031, when stored at recommended conditions.	
Quality control specifications	- SD082 has a diploid status for the sequences targeted by the reference probes and by the probes for the following genes: <i>PMS2</i>, <i>PMS2CL</i>, <i>SMN1</i> and <i>SMN2</i>, as present in SALSA® MLPA® Probemix P008-C1 PMS2, SALSA® MLPA® P021-B1 SMA and SALSA® MLPA® P060-B2 SMA Carrier. - Signal-to-size drop or sloping in the SD082 reactions should be similar (max 30% difference) as compared to arbitrarily chosen genomic samples known to be of sufficient quality for MLPA. - Concentration of SD082 is between 8.0 and 12.5 ng/μl.	Test result
		PASS

This lot was certified by MRC Holland on 20 February 2026.

More information: www.mrcholland.com ; www.mrcholland.eu	
	MRC Holland BV; Willem Schoutenstraat 1 1057 DL, Amsterdam, The Netherlands
E-mail	info@mrcholland.com (information & technical questions) order@mrcholland.com (orders)
Phone	+31 888 657 200

This certificate is a declaration of analysis at the time of the manufacturing process. All assays were run in compliance with manufacturer's instructions for use.

Implemented changes in the COA
Version 01 – 29 May 2026 (05) - Not applicable, new document.